

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

LaserGas Monitor HCl

manufactured by:

NEO Monitors AS

Solheimveien 62A
1473 Lørenskog
Norway

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Emission
Monitoring Systems, Version 2, Revision 1 (April 2003)**

Certification Ranges :

| | |
|-----|---------------------------|
| HCl | 0 to 15 mg/m ³ |
| | 0 to 90 mg/m ³ |

| | |
|-------------------------|-------------------|
| Project No: | 674/0224 |
| Certificate No: | Sira MC 060086/00 |
| Initial Certification: | 20 November 2006 |
| This Certificate Issued | 20 November 2006 |
| Renewal Date: | 19 November 2011 |


Technical Director

MCERTS is operated on behalf of the Environment Agency by

Page 1 of 4

Sira Certification Service

12 Acorn Industrial Park, Crayford Road,
Crayford, Dartford, Kent DA1 4AL

Tel: +44 (0)1322 520500 Fax: +44 (0)1322 520501

